



## SCS Directory

Accreditation number: SCS 0065

International standard: ISO/IEC 17025:2017  
Swiss standard: SN EN ISO/IEC 17025:2018

ROTRONIC AG  
Calibration laboratory  
Grindelstrasse 6  
8303 Bassersdorf

Head: Raphael von Bergen  
Responsible for MS: Andreas Gähwiler  
Telephone: +41 44 838 12 21  
E-Mail: <mailto:scs@rotronic.ch>  
Internet: <http://www.rotronic.ch>  
Initial accreditation: 27.04.1995  
Current accreditation: 31.10.2019 to 30.10.2024  
Scope of accreditation see: [www.sas.admin.ch](http://www.sas.admin.ch)  
(Accredited bodies)

Site under the accreditation:

Einsteinstrasse 17 - 23  
D-76275 Ettlingen

### Scope of accreditation as of 31.10.2019

#### Calibration laboratory for temperature and humidity

##### Calibration and Measurement Capability (CMC)

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Capability $\pm$ <sup>1)</sup>	Remarks
<b>TEMPERATURE</b>	-10 °C ... < 0 °C	In air	0,25 °C	
	0 °C ... < 21 °C		0,15 °C	
	21 °C ... < 25 °C		0,10 °C	
	25 °C ... < 35 °C		0,15 °C	
	35 °C ... < 50 °C		0,20 °C	
	50 °C ... 70 °C		0,25 °C	
Resistive temperature device PT100	- 25 °C ... 125 °C	In bath	0,03 °C	
Probes with analogue output	- 25 °C ... 125 °C	In bath	0,05 °C	



**SCS Directory**

**Accreditation number: SCS 0065**

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Capability $\pm$ <sup>1)</sup>	Remarks	
<b>RELATIVE HUMIDITY</b>	0,5 % rF ... < 20 % rF	<b>AMBIENT TEMPERATURE:</b> (23 $\pm$ 2) °C	0,2 % rF		
	20 % rF ... < 40 % rF	(23 $\pm$ 2) °C	0,3 % rF		
	40 % rF ... < 65 % rF	(23 $\pm$ 2) °C	0,4 % rF		
	65 % rF ... < 85 % rF	(23 $\pm$ 2) °C	0,5 % rF		
	85 % rF ... 99 % rF	(23 $\pm$ 2) °C	0,6 % rF		
		<b>AIR CONDITIONED BOX ROOM</b>			
	10 % rF ... < 20 % rF	-10 °C ... < 0 °C	0,6 % rF		
	20 % rF ... < 40 % rF	-10 °C ... < 0 °C	0,9 % rF		
	40 % rF ... < 65 % rF	-10 °C ... < 0 °C	1,5 % rF		
	65 % rF ... 95 % rF	-10 °C ... < 0 °C	2,1 % rF		
	10 % rF ... < 20 % rF	0 °C ... < 10 °C	0,3 % rF		
	20 % rF ... < 40 % rF	0 °C ... < 10 °C	0,6 % rF		
	40 % rF ... < 65 % rF	0 °C ... < 10 °C	0,9 % rF		
	65 % rF ... 95 % rF	0 °C ... < 10 °C	1,2 % rF		
	10 % rF ... < 20 % rF	10 °C ... < 35 °C	0,2 % rF		
	20 % rF ... < 40 % rF	10 °C ... < 35 °C	0,5 % rF		
	40 % rF ... < 65 % rF	10 °C ... < 35 °C	0,7 % rF		
	65 % rF ... 95 % rF	10 °C ... < 35 °C	0,9 % rF		
	10 % rF ... < 20 % rF	35 °C ... < 50 °C	0,3 % rF		
	20 % rF ... < 40 % rF	35 °C ... < 50 °C	0,6 % rF		
40 % rF ... < 65 % rF	35 °C ... < 50 °C	0,8 % rF			
65 % rF ... 95 % rF	35 °C ... < 50 °C	1,0 % rF			
10 % rF ... < 20 % rF	50 °C ... 70 °C	0,4 % rF			
20 % rF ... < 40 % rF	50 °C ... 70 °C	0,6 % rF			
40 % rF ... < 65 % rF	50 °C ... 70 °C	1,0 % rF			
65 % rF ... 95 % rF	50 °C ... 70 °C	1,6 % rF			



**SCS Directory**

**Accreditation number: SCS 0065**

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Capability $\pm$ <sup>1)</sup>	Remarks
<b>HUMIDITY STANDARDS OF OUR OWN PRODUCTION</b>	0,15 % rF.. < 25 % rF	(23 $\pm$ 2) °C	0,3 % rF	Only Site Switzerland
	25 % rF ... < 40 % rF	(23 $\pm$ 2) °C	0,4 % rF	Only Site Switzerland
	40 % rF ... < 70 % rF	(23 $\pm$ 2) °C	0,6 % rF	Only Site Switzerland
	70 % rF ... < 85 % rF	(23 $\pm$ 2) °C	0,7 % rF	Only Site Switzerland
	85 % rF ... < 99 % rF	(23 $\pm$ 2) °C	0,8 % rF	Only Site Switzerland

\* / \* / \* / \* / \*