



## SCS Directory

Accreditation number: SCS 0065

International standard: ISO/IEC 17025:2017

Swiss standard: SN EN ISO/IEC 17025:2018

ROTRONIC AG  
Calibration laboratory  
Grindelstrasse 6  
8303 Bassersdorf

Head:

Raphael von Bergen

Responsible for MS:

Raphael von Bergen

Telephone:

+41 44 838 12 21

Site under the accreditation:

E-Mail:

[calibration@rotronic.ch](mailto:calibration@rotronic.ch)

Internet:

[www.rotronic.ch](http://www.rotronic.ch)

Einsteinstrasse 17 - 23  
D-76275 Ettlingen

Initial accreditation:

27.04.1995

Current accreditation:

31.10.2024 to 30.10.2029

Scope of accreditation see:

[www.sas.admin.ch](http://www.sas.admin.ch)  
(Accredited bodies)

### Scope of accreditation as of 31.10.2024

#### Calibration laboratory for temperature and relative humidity (rh)

##### Calibration and Measurement Capability (CMC)

Calibration and measurement capabilities below carried out at the headquarters and office, in the laboratory at ambient temperature: 21 °C ... 25 °C

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Capability $\pm$ <sup>1)</sup>	Remarks
<b>TEMPERATURE</b>	-10 °C ... < 0 °C		0,25 °C	In climatic chamber
	0 °C ... < 35 °C		0,15 °C	In climatic chamber
	35 °C ... < 50 °C		0,20 °C	In climatic chamber
	50 °C ... 70 °C		0,25 °C	In climatic chamber
	21 °C ... 25 °C		0,1 °C	In humidity generator
Resistive temperature device PT100	-25 °C ... 125 °C		0,03 °C	In temperature bath
Digital probe	-25 °C ... 125 °C		0,04 °C	In temperature bath



## SCS Directory

Accreditation number: SCS 0065

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Capability $\pm$ <sup>1)</sup>	Remarks
RELATIVE HUMIDITY	0,5 % rh ... < 20 % rh	21 °C ... 25 °C	0,2 % rh	In humidity generator
	20 % rh ... < 40 % rh	21 °C ... 25 °C	0,3 % rh	In humidity generator
	40 % rh ... < 65 % rh	21 °C ... 25 °C	0,4 % rh	In humidity generator
	65 % rh ... < 85 % rh	21 °C ... 25 °C	0,5 % rh	In humidity generator
	85 % rh ... 99 % rh	21 °C ... 25 °C	0,6 % rh	In humidity generator
	10 % rh ... < 20 % rh	-10 °C ... < 0 °C	0,6 % rh	In climatic chamber
	20 % rh ... < 40 % rh	-10 °C ... < 0 °C	0,9 % rh	In climatic chamber
	40 % rh ... < 65 % rh	-10 °C ... < 0 °C	1,5 % rh	In climatic chamber
	65 % rh ... 95 % rh	-10 °C ... < 0 °C	2,1 % rh	In climatic chamber
	10 % rh ... < 20 % rh	0 °C ... < 10 °C	0,5 % rh	In climatic chamber
	20 % rh ... < 40 % rh	0 °C ... < 10 °C	0,6 % rh	In climatic chamber
	40 % rh ... < 65 % rh	0 °C ... < 10 °C	0,9 % rh	In climatic chamber
	65 % rh ... 95 % rh	0 °C ... < 10 °C	1,2 % rh	In climatic chamber
	10 % rh ... < 20 % rh	10 °C ... < 35 °C	0,4 % rh	In climatic chamber
	20 % rh ... < 40 % rh	10 °C ... < 35 °C	0,5 % rh	In climatic chamber
	40 % rh ... < 65 % rh	10 °C ... < 35 °C	0,7 % rh	In climatic chamber
	65 % rh ... 95 % rh	10 °C ... < 35 °C	0,9 % rh	In climatic chamber
	10 % rh ... < 20 % rh	35 °C ... < 50 °C	0,5 % rh	In climatic chamber
	20 % rh ... < 40 % rh	35 °C ... < 50 °C	0,6 % rh	In climatic chamber
	40 % rh ... < 65 % rh	35 °C ... < 50 °C	0,8 % rh	In climatic chamber
65 % rh ... 95 % rh	35 °C ... < 50 °C	1,0 % rh	In climatic chamber	
10 % rh ... < 20 % rh	50 °C ... 70 °C	0,5 % rh	In climatic chamber	
20 % rh ... < 40 % rh	50 °C ... 70 °C	0,6 % rh	In climatic chamber	
40 % rh ... < 65 % rh	50 °C ... 70 °C	1,0 % rh	In climatic chamber	
65 % rh ... 95 % rh	50 °C ... 70 °C	1,6 % rh	In climatic chamber	



## SCS Directory

Accreditation number: SCS 0065

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Capability $\pm$ <sup>1)</sup>	Remarks
<b>HUMIDITY STANDARDS OF OUR OWN PRODUCTION</b>	0,15 % rh ... < 25 % rh	21 °C ... 25 °C	0,3 % rh	Site Switzerland
	25 % rh ... < 40 % rh	21 °C ... 25 °C	0,4 % rh	Site Switzerland
	40 % rh ... < 70 % rh	21 °C ... 25 °C	0,6 % rh	Site Switzerland
	70 % rh ... < 85 % rh	21 °C ... 25 °C	0,7 % rh	Site Switzerland
	85 % rh ... < 99 % rh	21 °C ... 25 °C	0,8 % rh	Site Switzerland

The calibration and measurement capabilities below are carried out on site, at the headquarters and in the office. In the laboratory and at the customer's premises at ambient temperature: 15 °C ...30°C.

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Capability $\pm$ <sup>1)</sup>	Remarks
Relative humidity and Temperature	10 % rh ... 90 % rh	10 °C ... 40 °C	1,3 % rh 0,3 °C	In temperature and humidity generator
<b>TEMPERATURE</b>	-90 °C ... < 0 °C		0,4 °C	In dry block calibrator
	0 °C ... 125 °C		0,25 °C	In dry block calibrator

In case of contradictions in the language versions of the directories, the German version shall apply.

\* / \* / \* / \* / \*